

# AVRASYA CEPHE-DOĞRAMA TEST VE TEKNOLOJİ MERKEZİ A.Ş.

Sayfa (Page): 11 / 11

Ek 3: Test Cihazı Kalibrasyon Raporu  
Annex 3: Calibration Document of Testing Equipments

AB-0926-T

19-089-PR01

03.2020

## Evidence of Performance

Calibration of a test rig to determine joint permeability, watertightness and wind load

No. 19-003074-PR01  
(Kal. B-M01-KA12-en-01)



Client	AVRASYA Cephe-Dograma Test ve Teknoloji Merkezi A.S. 2. Organize Sanayi Bölgesi Cumhuriyet Mah. 1. Cadde No:1 81600 Beyköy - Düzce Turkey
Product	Window and facade test rig
Designation	Machine number 12-000825, Year of construction 2012
ift- number	26263
Measuring interval	Air pressure: -10000...-40; +40...+10000 Pa Air flow rate: -820...-0.5; +0.5...+820 m³/h i.N. Deflection: 0...+90 mm Water flow rate: 2.5...275 l/min
Test wall	Clamping area: window facade width: 4000 mm 12000 mm height: 3800 mm 12500 mm
Special features	---

### Basis \*)

EN 13830:2003-09  
EN 14351-1:2006+A2:2016-09/ift-  
KAL 2495-2 KA  
Durchfluss Luft 2015-03  
ift-KAL 2991 KA Länge 2012-09  
ift-KAL 3231 KA  
Durchfluss Wasser 2013-05  
PTB-Richtlinie DKD-R 6-  
1 rev2 2014-03

\*) Corresponds to the national standard/s  
(e.g. DIN EN)

### Representation



### Instructions for use

This calibration record can be used as evidence of suitability of the test installations in accordance with the above basis.

### Result

#### The requirements are fulfilled:

**Pressure and flow system** according to EN 1026, EN 1027, EN 12211, EN 12153, EN 12155 and EN 12179

The specified exp. measurement uncertainty and accuracy<sup>\*)</sup> of  $\leq \pm 5\%$  of measured value is fulfilled.

**Displacement transducer** according to EN 12179 and EN 12211

The specified accuracy<sup>\*)</sup> of  $\leq \pm 0.1$  mm and  $\leq \pm 5\%$  of measured value is fulfilled.

**Water flow/volume for watertightness test** according to EN 1027 and EN 12155

The specified accuracy<sup>\*)</sup> of  $\leq \pm 10\%$  of measured value is fulfilled.

\*) Accuracy = Deviation + expanded measurement uncertainty



### Validity

The data and results given relate solely to the tested and described test equipment.

Date of calibration:  
07.08.2019

Recommended re-calibration:  
07.11.2020

### Notes on publication

The ift-Guidance Sheet "Conditions and Guidance for the Use of ift Test Documents" applies. The cover sheet can be used as abstract.

### Contents

The report contains a total of 1 page/s including the calibration certificates listed in the following Annex

- 1 Cal. of pressure transducer
- 2 Cal. of air flow sensor
- 3 Cal. of displacement transducer
- 4 Cal. of water flow sensor

ift Rosenheim

Christine Lux, Dipl.-Phys.  
Head  
Calibration Laboratory

Paul König  
Calibration Expert  
Calibration Laboratory

ift Rosenheim GmbH  
Theodor-Gieß-Str. 7-9  
D-83026 Rosenheim

Kontakt  
Tel. +49 8031 261-0  
Fax +49 8031 261-290  
www.ift-rosenheim.de

Prüfung und Kalibrierung – EN ISO/IEC 17025  
Inspektion – EN ISO/IEC 17020  
Zertifizierung Produkte – EN ISO/IEC 17065  
Zertifizierung Managementsysteme – EN ISO/IEC 17021

Notified Body 8757  
PTB-Zeile BAY 18

